

TECHNICAL DATA SHEET

STERILIZATION REEL made of DuPont™ Tyvek®

- PRODUCT CODE** : TY
- RELATED STANDARDS** : PMSSteripack Sterilization Reels made of DuPont™ Tyvek® are manufactured according to EN 868-5 and ISO 11607 standards. The indicator used in the pouch production is classified as Class 1 Type Indicator according to the ISO 11140-1 standard. PMSSteripack Tyvek Reels meet the requirements of (EU) 2017/745 Medical Device Regulation.
- INTENDED USE** : The Sterilization Reels made of DuPont™ Tyvek® are used to sterilize Temperature-sensitive medical devices. It is preferred at Low Temperature Sterilization regarding to its porous structure. Reels meet the Medical Industry's requirement for high quality, hygiene and safety levels. PMSSteripack Reels are used for the packing of the Medical Devices before sterilization and it maintains the sterility of the products after sterilization.
- OPERATION CONDITIONS** : PMSSteripack Sterilization Reels made of DuPont™ Tyvek® are used for ethylene oxide, gas plazma (VH2O2) sterilization. The sterilization conditons should be determined by the end user regarding to material to be sterilized.
- SPECIFICATIONS** :

	PROPERTIES	UNIT	TYPICAL VALUE
TYVEK®	DELAMINATION	N/2.54 cm	2,7
	BASIC WEIGHT	g/m ²	60
	POROSITY	sec/100 cc	22
	OPACITY	%	94
	THICKNESS	µm	155
	SPENCER PUNCTURE	J/m	5254
	MULLEN BURST	kPa	925
	HYDROSTATIC HEAD	cm H2O	145
	ELMENDORF TEAR, CD	mN	3694
	ELMENDORF TEAR, MD	mN	2803
	TENSILE STRENGTH,CD	N/2.54 cm	157
	TENSILE STRENGTH,MD	N/2.54 cm	156
	FREE FROM LEAD AND HEAVY METALS AND TOXIC MATERIALS		



PMSSteripack®

STERILIZATION REEL made of DuPont™ Tyvek®

LAMINATED FILM	PROPERTIES	UNIT	STANDARD OR TEST METHOD	TYPICAL VALUE		
	THICKNESS	µm	ISO 534	LAM.FILM	64 ± 5	
				PE	50 ± 5	
				PET	12 ± 2	
	THERMAL SEAL	°C	ASTM F 88	110 ± 5		
	TENSILE	N/15 mm	MD CD	ASTM D 882-12	min. 26	
					min. 26	
	TEAR	mN	MD CD	ASTM D 1922	min. 90	
min. 90						
HAZE	%	ASTM D 1003	max. 17			
FREE FROM LEAD AND HEAVY METALS AND TOXIC MATERIALS						

PRODUCT SPECIFICATIONS

REEL	PROPERTIES	UNIT	STANDARD OR TEST METHOD	TYPICAL VALUE		
	SEAL STRENGTH	cm	ASTM F 88	< 25	≥ 25	
		Edge Seal (N/15 mm)		3,0 ± 0,5	4,0 ± 0,75	
	BUBBLE TEST		ASTM F 2096-04	None		
	PINHOLE DETERMINATION		EN 868-5 Annex C	None		
	DIMENSION CONTROL	cm or mm	ASTM F 2203-02	Required dimensions		
	LEAKAGE TEST		ASTM F 1929-12	None		
	PEEL DIRECTION		EN 868-5 Annex E	Must not break the particle		
INDICATOR CONTROL		ISO 11140-1	Must return to the specified color			

INDICATOR	STERILIZATION METHOD	BEFORE STERILIZATION	AFTER STERILIZATION
		VH2O2	Red

PACKAGING












: PMSSteripack TYVEK® reels are packaged as below.

Product Code	Pieces in Inner Packaging	Pieces in carton
05 cm	6	12
7,5 cm	4	8
10 cm	3	6
15 cm	2	4
20 cm	2	4
25 cm	1	2
30 cm	1	2
35 cm	1	2
40 cm	1	2



PMSSteripack®

BOX LABEL

PMS Steripack®		PMS SteriSeal™		PMS SteriTest™	
DESCRIPTION TANIM BESCHREIBUNG DEFINITION HEPITPAQH توضیو		SIZE EBAT GRÖÖE MESURE MEPEOOS الحجم			
CODE KOD CODE CODE ΚΩΔΙΚΟΣ رڈڈر		QUANTITY MIKTAR MENGE QUANTITE HOΣOYHTA الكبفة			
  LOT		UDI			
	 35 °C 85 °F				
	PMS Tıbbi Cihazlar Teknolojisi San. ve Tic. A.Ş. Karatuvır Mh. Serbest Bölge 11. Cadde No: 46, Adres No: 3191520954, Akdeniz 33020 Mersin / Türkiye T: +90 324 238 7042 F: +90 324 238 6549 www.pmsmedikal.com pms@pmsmedikal.com MADE IN TURKEY				
					EN 868-5



PMS Steripack®

PMS Tıbbi Cihazlar Teknolojisi San. ve Tic. A.Ş
pms@pmsmedikal.com